



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Sonic Clinical Trials**, using **Bio-Rad D-10** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: D-100	Calibrator Lot: AA90386, AA90387	Column Lot: R70468Z
Reagent Lot: AA00092, AA00093, AA00029	Calibrator Assigned Values: 5.4%, 10.2%	

Date of Certification: October 1, 2020

Certification Expires: October 1, 2021

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor