



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Sonic Clinical Trials*, using *Bio-Rad D-10* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

| | | |
|--|---|-------------------------------|
| Instrument: D-10 | Calibrator Lot: AA90386, AA90387 | Column Lot: R90055P |
| Reagent Lot: AA10090, AA10091, AA10092 | Calibrator Assigned Values: 5.4%, 10.3% | |

Date of Certification: October 1, 2021

Certification Expires: October 1, 2022

David Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor